

90-2167: Human Hif1 Alpha Pre Coated ELISA Kit

Application : ELISA

Description

This assay employs an antibody specific for Anti-human HIF1 α coated on a 96-well plate. Standards and samples are pipetted into the wells and HIF1 α present in a sample is bound to the wells by the immobilized antibody. The wells are washed and biotinylated anti-human HIF1 α antibody is added. After washing away unbound biotinylated antibody, HRP-conjugated streptavidin is pipetted to the wells. The wells are again washed, a TMB substrate solution is added to the wells and color develops in proportion to the amount of HIF1 α bound. The Stop Solution changes the color from blue to yellow, and the intensity of the color is measured at 450 nm.

Kit Components

| Item | Specifications | Storage |
|---|----------------|----------------------------|
| 96 well Strip ELISA Plate | 8 × 12 well | 4°C |
| Lyophilized Standard | 2 vials | -20°C |
| Sample and Standard Dilution Buffer | 20 ml | 4°C |
| Biotinylated Detection Antibody for h HIF1 α | 120 μ l | 4°C (Protected from light) |
| Antibody Dilution Buffer | 10 ml | 4°C |
| HRP Conjugated Streptavidin | 120 μ l | 4°C (Protected from light) |
| Streptavidin Dilution Buffer | 10 ml | 4°C |
| TMB Substrate | 10 ml | 4°C |
| Stop Solution | 10 ml | 4°C |
| 25X Wash Buffer | 30 ml | 4°C |
| Plate Sealer | 5 pieces | |
| Product Manual | 1 | |

Product Info

Amount : 1 x 96 Tests
Content : 1 x 96 well Format (96 tests)
Storage condition : Please refer to the Manual

Application Note

This immunoassay kit allows for the in vitro quantitative determination of Human HIF1 α concentrations in serum, plasma and other biological fluids.

Detection Range: 0.156 - 10 ng/ml

Sensitivity: < 0.094 ng/ml




Fig-1: Human HIF1 α Standard Curve .