

## 90-2393: Human PD-L1 Pre-Coated ELISA Kit

**Application :** ELISA

### Description

This assay employs an antibody specific for Anti-human PD-L1 coated on a 96-well plate. Standards and samples are pipetted into the wells and PD-L1 present in a sample is bound to the wells by the immobilized antibody. The wells are washed and biotinylated anti-human PD-L1 antibody is added. After washing away unbound biotinylated antibody, HRP-conjugated streptavidin is pipetted to the wells. The wells are again washed, a TMB substrate solution is added to the wells and color develops in proportion to the amount of PD-L1 bound. The Stop Solution changes the color from blue to yellow, and the intensity of the color is measured at 450 nm.

### Kit Components

Item	Specifications	Storage
96 well Strip ELISA Plate	8 × 12 well	4°C
Lyophilized Standard	2 vials	-20°C
Sample and Standard Dilution Buffer	20 ml	4°C
Biotinylated Detection Antibody for hPDL1	120 µl	4°C (Protected from light)
Antibody Dilution Buffer	10 ml	4°C
HRP Conjugated Streptavidin	120 µl	4°C (Protected from light)
Streptavidin Dilution Buffer	10 ml	4°C
TMB Substrate	10 ml	4°C
Stop Solution	10 ml	4°C
25X Wash Buffer	30 ml	4°C
Plate Sealer	5 pieces	
Product Manual	1	

### Product Info

**Amount :** 1 kit— 96 Tests  
**Content :** 1 kit— 96 well Format (96 tests)  
**Storage condition :** Please refer to the Manual

## Application Note

This immunoassay kit allows for the in vitro quantitative determination of Human PD-L1 concentrations in serum, plasma and other biological fluids. This kit is for Research use only. Not for use in diagnostic/therapeutics procedures.

**Detection Range:** 0.156 - 10 ng/ml

**Sensitivity:** < 94 pg/ml




Fig-1: Human PD-L1 Standard Curve .