

30-2848: Anti-Human CD2 FITC MAb (Clone: TS1/8)

Clonality :	Monoclonal
Clone Name :	TS1/8
Application :	FACS
Reactivity :	Human
Conjugate :	FITC
Gene :	CD2
Gene ID :	914
Uniprot ID :	P06729
Alternative Name :	T11, LFA-2, SRBC
Isotype :	Mouse IgG1 kappa
Immunogen Information :	Cytotoxic T lymphocytes

Description

Specificity: The mouse monoclonal antibody TS1/8 recognizes an extracellular epitope of CD2, a 50 kDa glycoprotein present on the human peripheral blood T lymphocytes and NK cells; also expressed by all thymocytes.

CD2 belongs to T lymphocyte glycoproteins of immunoglobulin superfamily. Its interaction with CD58 stabilizes adhesion between T cells and antigen presenting or target cells. Relatively low affinity of CD2 to CD58 (as measured in solution) is compensated within the two-dimensional cell-cell interface to provide tight adhesion. Moreover, T cell activation induces increased CD2 expression and its lateral mobility, making easier contact between CD2 and CD58. Subsequently, T cell activation causes fixation of CD58-CD2 at sites of cell-cell contact, thereby strengthening intercellular adhesion. CD2 deficiency reduces intestinal inflammation and helps to control infection.

Product Info

Amount :	100 tests
Purification :	Purified antibody is conjugated with fluorescein isothiocyanate (FITC) under optimum conditions and unconjugated antibody and free fluorochrome are removed by size-exclusion chromatography.
Content :	Stabilizing phosphate buffered saline (PBS), pH 7.4, 15 mM sodium azide
Storage condition :	Store at 2-8°C. Protect from prolonged exposure to light. Do not freeze.

Application Note

Flow cytometry: The reagent is designed for analysis of human blood cells using 4 µl reagent / 100 µl of whole blood or 10⁶ cells in a suspension. The content of a vial (0.4 ml) is sufficient for 100 tests.




Figure 1: Flow cytometry surface staining pattern of human peripheral whole blood stained using anti-human CD2 (TS1/8) FITC antibody (4 μ l reagent / 100 μ l of peripheral whole blood).




Figure 2: Separation of human CD2 positive lymphocytes (red-filled) from CD2 negative lymphocytes (black-dashed) in flow cytometry analysis (surface staining) of human peripheral whole blood stained using anti-human CD2 (TS1/8) FITC antibody (4 μ l reagent / 100 μ l of peripheral whole blood).